

Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 12/30/2014

lms
T/N

Instrument Serial Number: 010506

As of 30-Dec-14

Start Date	End Date	Agency Name
29-Dec-14		DFS Central Lab
26-Aug-14	29-Dec-14	Fairfax County PD
05-Aug-14	26-Aug-14	DFS Central Lab
31-Jan-14	05-Aug-14	Powhatan County Sheriff's Office
07-Sep-11	31-Jan-14	DFS Central Lab
04-Aug-11	07-Sep-11	Richmond PD
06-Jan-09	04-Aug-11	DFS Central Lab
20-Nov-08	06-Jan-09	Waynesboro PD
19-May-08	20-Nov-08	DFS Central Lab

LMS
T/V

INTOX EC/IR II

Quality Assurance Worksheet

Instrument Serial Number 010506 Worksheet Start Date 12/29/2014

Location Fairfax County Adult Detention Center

Address 10520 Judicial Dr., Fairfax, VA 22030

DFS Technician Heather Stanton License No. 18910

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 755 Reference Barometer (mm HG) 756

Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/28/2015

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.) AG332301-05

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:

Notes:

Unable to perform instrument test and measurement assurance checks due to instrument being disabled.

Instrument Serial Number

010506

Certification Date

☐ Calibrated☐ Certified☐ Measurement Assurance Check☐ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath W. Storch

Date

12/30/14

Issuing Analyst

J. J. P.

Date

12/30/14

JPP

Department of Forensic Science
Breath Alcohol Section
Trouble Call Log
12/28/2014 To 12/28/2014

WMS
T.L. ✓

Instrument Serial Number: 010506

30-Dec-14

Date: 28-Dec-14 Time: 8:55 PM Remote: True Tech: HMS (18910) Location: Fairfax County PD

Indication: DISABLED

Problem: INSTRUMENT DISABLED DUE TO OUT OF TOLERANCE MESSEGES.

Technician Response: SITE VISIT CONDUCTED.

Resolution REMOVED TO DFS-Central
